



User Manual

Sunset Compressor Nebulizer

NEB100



Instructions for Use

The Sunset Compressor Nebulizer is an AC-powered air nebulizer compressor system intended to provide a source of compressed air for medical purposes for use in home healthcare. It is to be used with a nebulizer kit to produce aerosol particles of medication for respiratory therapy for both children and adults.

Important: Read these instructions carefully before using this device.

Caution

1. Use the device only for its intended purpose as described in these instructions for use as a compressor nebulizer system for aerosol therapy. Follow the instructions of your doctor. Any other form of use constitutes improper use and may be dangerous. The manufacturer cannot be held liable for any damage caused by improper, incorrect or unreasonable use, or if the unit is connected to electrical installations not complying with safe regulations.
2. Keep this manual for future reference.
3. Do not operate the unit in the presence of any anesthetic, inflammable mixtures or oxygen.
4. Correct device operation can be affected by electromagnetic interference. If this device interferes with another electrical apparatus, remove it and plug it into a different electrical outlet.
5. In case of failure and/or malfunction, do not open the compressor housing as this is not serviceable product.
6. The proper function of the device can be compromised if original replacement parts (e.g. filter) are not used.
7. When using any electrical appliance, certain important safety measures must always be observed, including the following:
 - Use only original accessories and components;
 - Never submerge the unit in water, as the device is not protected against water penetration;
 - Never touch the unit if hands are wet or moist;
 - Do not leave the unit outdoors;

- Place on stable and horizontal surface when operating;
 - Make sure the air vent openings area not obstructed;
 - Do not allow unsupervised children or physically challenged individuals to use the unit;
 - Do not turn the unit OFF simply by pulling out the plug from electrical outlet
8. Make sure that the electrical rating shown on the rating plate on the bottom of the unit corresponds to you main voltage and frequency before plugging in the device.
 9. Do not use any electrical adapter including extension cords.
 10. Do not leave the unit plugged in when not in use. Unplug the device from the electrical outlet when not in operation.
 11. Follow the manufacturer's instructions for installing the device. The manufacturer cannot be held responsible for any damage caused by incorrect installation.
 12. The power supply cord cannot be replaced by the user.
 13. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
 14. Before performing any cleaning operation or filter replacement, disconnect the device from the main supply by unplugging it or by turning the main supply switch OFF.
 15. Some parts of the unit are small enough to be swallowed by children. Do not let unattended children have access to the unit.
 16. If you decide not to use the device, it is recommended to dispose of it according to local regulations.
 17. Remember to:
 - Use this device only with medications prescribed by your doctor;
 - Use the accessories recommended by your doctor for therapy;
 18. Verify that the medications prescribed by your doctor can be used with common compressor nebulizer systems for aerosol therapy.

Instructions for Use with T-Piece Kit

 Examine the unit before the first use in order to check for possible shipping damage. When taking a treatment sit upright and relaxed.

1. After unpacking the device, check the compressor and accessories for visible damage or defects such as cracks in the plastic housing, which could expose electrical components. Contact your retailer to report any damaged product.
 2. Before using the device, proceed with the cleaning operations as described in the cleaning section.
 3. Attach the supply tube to the compressor and to the bottom of the nebulizer cup.
 4. Open the nebulizer cup by pushing down on the cap and twisting counter-clockwise.
 5. Add prescribed medication into nebulizer cup then turn top on until secure.
 6. Attach T-piece to the nebulizer bottle.
 7. Attach mouthpiece and corrugated tube to T-piece.
 8. Connect the 7ft tube to the nebulizer bottle and the nebulizer machine.
 9. Ensure connections are secure before turning on the nebulizer machine.
 10. Plug the compressor into the electrical outlet, making sure the electrical rating shown on the rating plate on the bottom of the unit corresponds to your main voltage.
 11. To start treatment, switch the compressor on by pressing the ON/OFF switch onto the "I" position.
- ATTENTION: This device rated for intermittent use 30 minutes ON/ 30 minutes OFF.
12. When mist is visible, begin inhaling with the mouthpiece.
 13. When the treatment has been completed, switch the unit OFF by setting the switch position to "O" and disconnect the plug from the electrical outlet.
 14. Wash the nebulizer and accessories as described in the cleaning section.

Cleaning

Cleaning of the compressor

Use a clean dry cloth and mild detergents to clean the outside of the unit.

ATTENTION: During cleaning, make sure the internal parts of the unit are not in contact with liquids and the power cord is not connected.

Cleaning and disinfection of the accessories

After each use, disassemble T-piece kit into three sections.

Wash in warm soapy water and rinse well. Wipe with a paper towel and allow to air dry.

 Do not boil or autoclave angled mouthpieces, air tubes or masks

 It is important that the cleaning instructions are carefully followed to ensure optimal operation.

Replacement

The nebulizer T-piece kit should be replaced after a long period of inactivity, or if it is obstructed by dry medication, dust etc. With normal use, the included nebulizer T-piece kit should last 1 month.

Replacement of the filter

The filter is to be replaced after approximately 30 hours of use, or when it appears dirty. Open the filter holder to replace the filter with a new one. Use Sunset BF018 Nebulizer Filters.

 Do not use the device without the filter.

Troubleshooting

The device does not turn on:

1. Make sure the plug is firmly inserted into the electrical outlet.
2. Make sure the device has been operating within operating limits indicated in this instructional manual (30 minutes ON/ 30 minutes OFF)

The device does not nebulize, or the nebulization is weak:

1. Make sure the ends of the air tube are fitted tightly onto the main unit and onto the nebulizer.
2. Check whether the nebulizer mask or mouthpiece kit cup chamber is empty of medication or overfilled with medication.
3. Check whether the nebulizer nozzle is obstructed.
4. Check whether the baffle is inserted properly.

Maintenance and repair

Never open the device. There are no serviceable parts within the unit. The compressor requires no lubrication or maintenance.

